



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Calibration Field/ Quantity/ Property	Measurand / Equipment	Measurement Range	Expanded Measurement Uncertainty (k=2)	Calibration Method (Standard/ Internal Procedure)	Permanent lab (P)/Client-site
Measurement of force, weight and pressure	Pneumatic hand Pump With Gauge	- 80 kpa to 4 Mpa	1 kpa	(AMC-CP-P-01) Based on DKD-R 6-1,DKD-R 6-2, EURAMET CG 17	P
Measurement of force, weight and pressure	hydraulic hand Pump With Gauge	0 kpa to 70 Mpa	23 kpa	(AMC-CP-P-01) Based on DKD-R 6-1,DKD-R 6-2, EURAMET CG 17	P
Thermodynamics and temperature measurements	Freezer, Refrigerator	- 80 to 0 Degree	0.34 °C	(AMC-CP-T-01) Based on DKD-R5-7.	P
Thermodynamics and temperature measurements	Chillers, Oven , Refrigerator, auto Clave , Hot Cabinet , Dish Washer , Glass Washer	0 to 250 Degree	0.89 °C	(AMC-CP-T-01) Based on DKD-R5-7.	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Thermodynamics and temperature measurements	RTD Thermometer/Thermocouple	- 40°C to 200°C	0.4 °C	(AMC-CP-T-02) Based on Euramet CG-13 /V.3.0, Euramet CG-11 /V.2.0,	P
Thermodynamics and temperature measurements	RTD Thermometer/Thermocouple	200°C to 400°C	0.7 °C	(AMC-CP-T-02) Based on Euramet CG-13 /V.3.0, Euramet CG-11 /V.2.0,	P
Thermodynamics and temperature measurements	Thermo hygrometer	(10 to 50) °C / (20 to 90) % Rh	0.32 °C / 2.6% Rh	(AMC-CP-T-04) based on IEC 60068-3-11 : 2007.NABL 124, Euramet Cg 20,DKD-R-5-8	P
Thermodynamics and temperature measurements	Temperature Gauge, Pocket / Pen type Thermometer	- 40°C to 200°C	0.4 °C	(AMC-CP-T-05) Based on DKD-R5-1, Euramet CG-13 /V.3.0, NABL 124	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Thermodynamics and temperature measurements	Temperature Gauge, Pocket / Pen type Thermometer	200°C to 400°C	0.6 °C	(AMC-CP-T-05) Based on DKD-R5-1, Euramet CG-13 /V.3.0, NABL 124	P
Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	1 mg to 500 mg	0.2 mg	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V4	S
Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	>500 mg to 200 g	0.6 mg	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V5	S
Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	>200 g to 500 g	0.11 g	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V6	S

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	>500 g to 10 kg	0.12 g	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V7	S
Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	>10 kg to 20 kg	0.7 g	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V8	S
Measurement of force, weight and pressure	Electronic Balance / Weighing Scale	>20 kg to 300 kg	58 g	(AMC-CP-WB-01) Based on OIML R-76-1.2006 (E), OIML R-76-2.2006 (E), EUROMET CG 18 V9	S
Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<1mV to 200mV	13 μ V	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<200 mV to 2 V	1.1 mV	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<2V to 10 V	2.2 mV	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<10 V to 100 V	0.01 V	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<100 V to 200 V	0.03 V	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Voltage	<200 V to 1000 V	0.12 V	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Voltage	10 mV to 200 mV @ 50 Hz	0.12 mv	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	AC Voltage	<200 mV to 20 V @ 50 Hz	11 mV	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Voltage	<20 V to 200 V @ 50 Hz	0.09 V	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Voltage	<200 V to 1000 V @ 50 Hz	0.5 V	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Current	<1 mA to 2 mA	0.25 µA	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Current	<2 mA to 20 mA	2.0 µA	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Current	<20 mA to 200 mA	0.05 mA	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	DC Current	<200 mA to 2000 mA	1.1 mA	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	DC Current	<2 A to 20 A	0.02 A	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current	0.1 mA to 2 mA @ 50 Hz	2.0 µA	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current	< 2 mA to 20 mA @ 50 Hz	12 µA	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current	< 20 mA to 200 mA @ 50 Hz	0.19 mA	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current	< 200 mA to 2000 mA @ 50 Hz	3.1 mA	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	AC Current	< 2 A to 20 A @ 50 Hz	0.02 A	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Earth resistance tester	1 Ω to 10 Ω	5.1 m Ω	(AMC-UB-EL-03) Based on IS 11994: 1986 Euramet Cg-15. Ver.03	P
Electricity. Magnetism. Electrical and magnetic measurements	Earth resistance tester	<10 Ω to 100 Ω	26 m Ω	(AMC-UB-EL-03) Based on IS 11994: 1986 Euramet Cg-15. Ver.03	P
Electricity. Magnetism. Electrical and magnetic measurements	Earth resistance tester	<100 Ω to 10 K Ω	2.4 Ω	(AMC-UB-EL-03) Based on IS 11994: 1986 Euramet Cg-15. Ver.03	P
Electricity. Magnetism. Electrical and magnetic measurements	Earth resistance tester	< 10 K Ω to 1 M Ω	0.15 K Ω	(AMC-UB-EL-03) Based on IS 11994: 1986 Euramet Cg-15. Ver.03	P
Electricity. Magnetism. Electrical and magnetic measurements	Earth resistance tester	<1 M Ω to 10 M Ω	5.3 K Ω	(AMC-UB-EL-03) Based on IS 11994: 1986 Euramet Cg-15. Ver.03	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	AC Current(Clamp)	< 10 A to 100 A @ 50 Hz	0.38 A	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current(Clamp)	100 A to 500 A @ 50 Hz	2.0 A	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current(Clamp)	500 A to 900 A @ 50 Hz	5.0 A	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	AC Current (Clamp)	900 A to 1000A @ 50 Hz	10.0 A	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	<100 Ω to 1000 Ω	1.2 Ω	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	< 1 KΩ to 10 KΩ	0.01 kΩ	(AMC-UB-EL- 01) Based on EURAMET/cg- 15/v.3.0	P

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.



Accreditation Scope

ACCUVER AUTOMATION MECHANICAL AND PHYSICAL TESTS CO - L.L.C, NAL 234
Calibration Laboratory, (ISO/IEC 17025:2017)

WAREHOUSE – 5, PLOT NO: 27A14, ICAD III, ABU DHABI INDUSTRIAL CITY, ABU DHABI, Abu Dhabi,
United Arab Emirates

Issue Date: 16/05/2025

Expiry Date: 06/05/2028

Issue No: 2

Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	< 10 K Ω to 100 K Ω	0.11 k Ω	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	<0.1 M Ω to 1 M Ω	0.01 M Ω	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	<1 M Ω to 10 M Ω	0.11 M Ω	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
Electricity. Magnetism. Electrical and magnetic measurements	Resistance(Measurement)	<10 M Ω to 100 M Ω	1.2 M Ω	(AMC-UB-EL-01) Based on EURAMET/cg-15/v.3.0	P
END					

This is an electronic certificate and does not require stamp and signature. Certificate will be invalid in case of any modification.



هذه الشهادة صدرت إلكترونياً ولا تحتاج لختم أو توقيع، أي كشط أو تغيير في هذه الشهادة يلغيها.